


राष्ट्रीय अंटार्कटिक एवं
समुद्री अनुसंधान केन्द्र
पृथ्वी विज्ञान मंत्रालय
भारत सरकार)
हेड लैण्ड सडा, वास्को डा गामा
गोवा- ४०३ ८०४ भारत



**NATIONAL CENTRE FOR
ANTARCTIC & OCEAN RESEARCH**

Ministry of Earth Sciences
Government of India)
Headland Sada, Vasco da Gama
Goa - 403 804 - INDIA

Tel/Fax No.: 0091 0832 252573

Email:warlu62@ncaor.gov.in

PROCUREMENT SECTION

Foreign Enquiry

M/s.

Ref. No.:

HSS -089

Date:

11.08.2016

Due Date:

29.08.2016

Dear Sir,

You are requested to send your offer for the following items as per the terms and conditions mentioned below.

SL. NO.	Description	Quantity
01.	Cooling Incubator (Specification as per Annexure I)	01

Your offer should contain the following information:

1. Submit your quotation on F.O.B/CIF Goa.
2. Delivery Period:
3. Validity of quotation:
4. Quantity/Trade discounts, if any. :
5. Guarantee/Warranty of the product:
6. Taxes applicable (VAT if any):
7. Specify Brand/Make/Model of the item:
8. Enclose brochure/leaflet – Specification:
9. Mode of payment acceptable to you: No advance payment will be made. Payment will be by Telegraphic Transfer immediately after dispatch of material.

Director, NCAOR reserves the right to reject any quotation received without assigning any reasons.

Yours faithfully

Sd/-

**Executive (Procurement)
For and behalf of Director, NCAOR**

SPECIFICATION FOR COOLING INCUBATOR.

FEATURES	SPECIFICATIONS
Type of incubator	Cooling incubator
Internal chamber volume	100-200 litres
Temperature range	Temperature range should be between $0^{\circ}\text{C}\pm 1^{\circ}$ to $80^{\circ}\text{C}\pm 10^{\circ}\text{C}$.
Internal chamber	Interior should be made up of stainless steel with at least four to five racks.
Closable port	The instrument should have closable port at the side for introducing pipelines/wires for connection.
Certification	CE certification should be provided for the quoted model.
Accessories	Stainless steel racks (4 numbers)
Warranty	One Year

TECHNICAL COMPLIANCE STATEMENT FOR SUPPLY OF COOLING INCUBATOR

SR. NO.	TECHNICAL COMPLIANCE STATEMENT FOR SUPPLY OF COOLING INCUBATOR		COMPLIED/ NOT COMPLIED	EXTRA FEATURES
1.	Type of incubator	Cooling incubator		
	Internal chamber volume	100-200 litres		
	Temperature range	Temperature range should be between 0°C±1° to 80°C±10°C.		
	Internal chamber	Interior should be made up of stainless steel with at least four to five racks.		
	Closable port	The instrument should have closable port at the side for introducing pipelines/wires for connection.		
	Certification	CE certification should be provided for the quoted model.		
	Accessories	Stainless steel racks (4 numbers)		
	Warranty	One Year		