


राष्ट्रीय अंटार्कटिक एवं
समुद्री अनुसंधान केन्द्र
पृथ्वी विज्ञान मंत्रालय
भारत सरकार)
हेड लैण्ड सडा, वास्को डा गामा
गोवा- ४०३ ८०४ भारत



**NATIONAL CENTRE FOR
ANTARCTIC & OCEAN RESEARCH**

Ministry of Earth Sciences
Government of India)
Headland Sada, Vasco da Gama
Goa - 403 804 - INDIA

Tel/Fax No.: 0091 0832 2525573

Email:warlu62@ncaor.gov.in

PROCUREMENT SECTION

Foreign Enquiry

M/s.

Ref. No.:

HSS -090

Date:

11.08.2016

Due Date:

29.08.2016

Dear Sir,

You are requested to send your offer for the following items as per the terms and conditions mentioned below.

SL. NO.	Description	Quantity
01.	Autoclave (Specification as per Annexure I)	01

Your offer should contain the following information:

1. Submit your quotation on F.O.B/CIF Goa.
2. Delivery Period:
3. Validity of quotation:
4. Quantity/Trade discounts, if any. :
5. Guarantee/Warranty of the product:
6. Taxes applicable (VAT if any):
7. Specify Brand/Make/Model of the item:
8. Enclose brochure/leaflet – Specification:
9. Mode of payment acceptable to you: No advance payment will be made. Payment will be by Telegraphic Transfer immediately after dispatch of material.

Director, NCAOR reserves the right to reject any quotation received without assigning any reasons.

Yours faithfully

Sd/-
Executive (Procurement)
For and behalf of Director, NCAOR

SPECIFICATION OF AUTOCLAVE.

FEATURES	SPECIFICATIONS
Type of Model	Floor model
Operating temperature	100-121° C or better
Operating pressure	10- 125 kpa or better
Lid opening/closing	Lid opening/closing should be manual
Internal chamber capacity	20-40 litre
Time display	Digital, microprocessor controlled
Temperature display	Digital, microprocessor controlled
Pressure gauge display	Analog
Sterilization, warming and heating program	Should have separate programmes for sterilization, warming and heating modes to keep microbiological media at desired physical state.
Safety features	Water level sensor, Over-heat protection, Over-pressure protection Empty heating protection Safety valve Incomplete sterilization detection
Power supply	Single -phase 120V AC 5-15A
Weight	40-50 kg
Transport	Instrument should have casters with stopper for easy transport
Accessories	Stainless basket (1 number) Bottom plate (1 number)
Certification	CE/ISO 9001 certification should be provided for the quoted model.
Warranty	1 year.

TECHNICAL COMPLIANCE STATEMENT FOR SUPPLY OF AUTOCLAVE

SR. NO.	TECHNICAL COMPLIANCE STATEMENT FOR SUPPLY OF AUTOCLAVE		COMPLIED/ NOT COMPLIED	EXTRA FEATURES
1.	Type of Model	Floor model		
	Operating temperature	100-121° C or better		
	Operating pressure	10- 125 kpa or better		
	Lid opening/closing	Lid opening/closing should be manual		
	Internal chamber capacity	20-40 litre		
	Time display	Digital, microprocessor controlled		
	Temperature display	Digital, microprocessor controlled		
	Pressure gauge display	Analog		
	Sterilization, warming and heating program	Should have separate programmes for sterilization, warming and heating modes to keep microbiological media at desired physical state.		
	Safety features	Water level sensor, Over-heat protection, Over-pressure protection Safety valve Incomplete sterilization detection		
	Power supply	Single –phase 120 V AC 5-15A		
	Weight	40-50 kg		
	Transport	Instrument should have casters with stopper for easy transport		
	Accessories	Stainless basket (1 number) Bottom plate (1 number)		
	Certification	CE/ISO 9001 certification should be provided for the quoted model.		
	Warranty	1 year.		