



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2023/B/4207611
Dated/दिनांक : 23-11-2023

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	14-12-2023 11:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	14-12-2023 11:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	60 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Earth Sciences
Department Name/विभाग का नाम	Department Of Earth Sciences
Organisation Name/संगठन का नाम	National Centre For Antarctic And Ocean Research (ncaor)
Office Name/कार्यालय का नाम	Goa
Total Quantity/कुल मात्रा	7
Item Category/मद केटेगरी	Test Sieves as per IS 460 (Part 1 and Part 2) (Q3)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Primary product category	Test Sieves as per IS 460 (Part 1 and Part 2)
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	4 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation

EMD Detail/ईएमडी विवरण

Required/आवश्यकता	No
-------------------	----

ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता	No
-------------------	----

Splitting/विभाजन

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

2. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)**Technical Specifications/तकनीकी विशिष्टियाँ**

[* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Standard Parameters	Conformity to standard of product	Woven wire cloth test sieves as per IS:460 (Part 1) - 1985
	BIS marked	Yes, No Or higher
Performance Parameters	Shape of Test sieves	Round
	Nominal size of test sieves (in mm)	200
	Material of wire cloth	Stainless Steel
	Material of perforated plate	NA for wire cloth
	Nominal Aperture size (width of aperture) of wire cloth	150 ?m
	Wire diameter of wire cloth sieves	140 ?m
	Wire diameter and Uniform Square Aperture size with tolerance of wire cloth table 2 in IS 460 (Part 1) 1985 standard provided	Yes
	Shape of Aperture provided in perforated plate	NA for wire cloth
	Thickness of perforated plate (in mm)	NA for wire cloth
	Nominal aperture size of holes with tolerance for perforated plate (in mm)	NA for wire cloth
	Frame Material of test sieves	Brass sheet
	Frame material thickness (in mm)	0.45
	Depth from top edge to sieving surface (in mm)	50
	Depth from bottom edge to sieving surface (in mm)	15
	Warranty in years	1
	ISO certification of manufacturer	Yes Or higher
Availability of test certificate from Government/ NABL/ ILAC accerdated lab to prove conformity of specifications	Yes Or higher	

Additional Specification Parameters - Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional Specification Parameter 1	Test sieve, brass: 4" dia., 150?m mesh size, with stainless steel mesh
Additional Specification Parameter 2	Sieve diameter-100 mm

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Pradeep Kumar Singh	403804, Director National Centre for Antarctic Ocean Research Headland Sada, Vasco-da-gama,	1	20

Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)**Technical Specifications/तकनीकी विशिष्टियाँ**

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Standard Parameters	Conformity to standard of product	Woven wire cloth test sieves as per IS:460 (Part 1) - 1985
	BIS marked	Yes, No Or higher
Performance Parameters	Shape of Test sieves	Round
	Nominal size of test sieves (in mm)	200
	Material of wire cloth	Stainless Steel
	Material of perforated plate	NA for wire cloth
	Nominal Aperture size (width of aperture) of wire cloth	212 ?m
	Wire diameter of wire cloth sieves	200 ?m

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Wire diameter and Uniform Square Aperture size with tolerance of wire cloth table 2 in IS 460 (Part 1) 1985 standard provided	Yes
	Shape of Aperture provided in perforated plate	NA for wire cloth
	Thickness of perforated plate (in mm)	NA for wire cloth
	Nominal aperture size of holes with tolerance for perforated plate (in mm)	NA for wire cloth
	Frame Material of test sieves	Brass sheet
	Frame material thickness (in mm)	0.45
	Depth from top edge to sieving surface (in mm)	50
	Depth from bottom edge to sieving surface (in mm)	15
	Warranty in years	1
	ISO certification of manufacturer	Yes Or higher
	Availability of test certificate from Government/ NABL/ ILAC accerdted lab to prove conformity of specifications	Yes Or higher

Additional Specification Parameters - Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional Specification Parameter 1	Test sieve, brass: 4" dia., 200µm mesh size, with stainless steel mesh
Additional Specification Parameter 2	Sieve diameter-100 mm

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Pradeep Kumar Singh	403804, Director National Centre for Antarctic Ocean Research Headland Sada, Vasco-da-gama,	1	20

Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Standard Parameters	Conformity to standard of product	Woven wire cloth test sieves as per IS:460 (Part 1) - 1985
	BIS marked	Yes, No Or higher
Performance Parameters	Shape of Test sieves	Round
	Nominal size of test sieves (in mm)	200
	Material of wire cloth	Stainless Steel
	Material of perforated plate	NA for wire cloth
	Nominal Aperture size (width of aperture) of wire cloth	250 ?m
	Wire diameter of wire cloth sieves	224 ?m
	Wire diameter and Uniform Square Aperture size with tolerance of wire cloth table 2 in IS 460 (Part 1) 1985 standard provided	Yes
	Shape of Aperture provided in perforated plate	NA for wire cloth
	Thickness of perforated plate (in mm)	NA for wire cloth
	Nominal aperture size of holes with tolerance for perforated plate (in mm)	NA for wire cloth
	Frame Material of test sieves	Brass sheet
Frame material thickness (in mm)	0.45	

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Depth from top edge to sieving surface (in mm)	50
	Depth from bottom edge to sieving surface (in mm)	15
	Warranty in years	1
	ISO certification of manufacturer	Yes Or higher
	Availibility of test certificate from Government/ NABL/ ILAC accerdited lab to prove conformity of specifications	Yes Or higher

Additional Specification Parameters - Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional Specification Parameter 1	Test sieve, brass: 4" dia., 250?m mesh size, with stainless steel mesh
Additional Specification Parameter 2	Sieve diameter-100 mm

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Pradeep Kumar Singh	403804,Director National Centre for Antarctic Ocean Research Headland Sada, Vasco-da-gama,	1	20

Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Standard Parameters	Conformity to standard of product	Perforated plate test sieves as per IS:460 (Part 2) - 1985
	BIS marked	Yes, No Or higher
Performance Parameters	Shape of Test sieves	Round
	Nominal size of test sieves (in mm)	200
	Material of wire cloth	Stainless Steel
	Material of perforated plate	Brass sheet
	Nominal Aperture size (width of aperture) of wire cloth	45 ?m
	Wire diameter of wire cloth sieves	45 ?m
	Wire diameter and Uniform Square Aperture size with tolerance of wire cloth table 2 in IS 460 (Part 1) 1985 standard provided	NA for Perforated plate
	Shape of Aperture provided in perforated plate	Square hole (C)
	Thickness of perforated plate (in mm)	NA for wire cloth
	Nominal aperture size of holes with tolerance for perforated plate (in mm)	NA for wire cloth
	Frame Material of test sieves	Brass sheet
	Frame material thickness (in mm)	0.45
	Depth from top edge to sieving surface (in mm)	50
	Depth from bottom edge to sieving surface (in mm)	15
	Warranty in years	1
	ISO certification of manufacturer	Yes Or higher
Availability of test certificate from Government/ NABL/ ILAC accerdated lab to prove conformity of specifications	Yes Or higher	

Additional Specification Parameters - Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional Specification Parameter	Test sieve, brass: 8" dia., 45?m mesh size, with stainless steel mesh

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Pradeep Kumar Singh	403804, Director National Centre for Antarctic Ocean Research Headland Sada, Vasco-da-gama,	1	20

Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)**Technical Specifications/तकनीकी विशिष्टियाँ**

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Standard Parameters	Conformity to standard of product	Woven wire cloth test sieves as per IS:460 (Part 1) - 1985
	BIS marked	Yes, No Or higher
Performance Parameters	Shape of Test sieves	Round
	Nominal size of test sieves (in mm)	200
	Material of wire cloth	Stainless Steel
	Material of perforated plate	NA for wire cloth
	Nominal Aperture size (width of aperture) of wire cloth	63 ?m
	Wire diameter of wire cloth sieves	63 ?m
	Wire diameter and Uniform Square Aperture size with tolerance of wire cloth table 2 in IS 460 (Part 1) 1985 standard provided	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Shape of Aperture provided in perforated plate	NA for wire cloth
	Thickness of perforated plate (in mm)	NA for wire cloth
	Nominal aperture size of holes with tolerance for perforated plate (in mm)	NA for wire cloth
	Frame Material of test sieves	Brass sheet
	Frame material thickness (in mm)	0.45
	Depth from top edge to sieving surface (in mm)	50
	Depth from bottom edge to sieving surface (in mm)	15
	Warranty in years	1
	ISO certification of manufacturer	Yes Or higher
	Availability of test certificate from Government/ NABL/ ILAC accerded lab to prove conformity of specifications	Yes Or higher

Additional Specification Parameters - Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional Specification Parameter	Test sieve, brass: 8" dia., 63µm mesh size, with stainless steel mesh

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Pradeep Kumar Singh	403804, Director National Centre for Antarctic Ocean Research Headland Sada, Vasco-da-gama,	1	20

Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Standard Parameters	Conformity to standard of product	Woven wire cloth test sieves as per IS:460 (Part 1) - 1985
	BIS marked	Yes, No Or higher
Performance Parameters	Shape of Test sieves	Round
	Nominal size of test sieves (in mm)	200
	Material of wire cloth	Stainless Steel
	Material of perforated plate	NA for wire cloth
	Nominal Aperture size (width of aperture) of wire cloth	300 ?m
	Wire diameter of wire cloth sieves	280 ?m
	Wire diameter and Uniform Square Aperture size with tolerance of wire cloth table 2 in IS 460 (Part 1) 1985 standard provided	Yes
	Shape of Aperture provided in perforated plate	NA for wire cloth
	Thickness of perforated plate (in mm)	NA for wire cloth
	Nominal aperture size of holes with tolerance for perforated plate (in mm)	NA for wire cloth
	Frame Material of test sieves	Brass sheet
Frame material thickness (in mm)	0.45	

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Depth from top edge to sieving surface (in mm)	50
	Depth from bottom edge to sieving surface (in mm)	15
	Warranty in years	1
	ISO certification of manufacturer	Yes Or higher
	Availability of test certificate from Government/ NABL/ ILAC accerded lab to prove conformity of specifications	Yes Or higher

Additional Specification Parameters - Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional Specification Parameter	Test sieve, brass: 8" dia., 300µm mesh size, with stainless steel mesh

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Pradeep Kumar Singh	403804, Director National Centre for Antarctic Ocean Research Headland Sada, Vasco-da-gama,	1	20

Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Standard Parameters	Conformity to standard of product	Woven wire cloth test sieves as per IS:460 (Part 1) - 1985
	BIS marked	Yes, No Or higher
Performance Parameters	Shape of Test sieves	Round
	Nominal size of test sieves (in mm)	300
	Material of wire cloth	Stainless Steel
	Material of perforated plate	NA for wire cloth
	Nominal Aperture size (width of aperture) of wire cloth	500 ?m
	Wire diameter of wire cloth sieves	500 ?m
	Wire diameter and Uniform Square Aperture size with tolerance of wire cloth table 2 in IS 460 (Part 1) 1985 standard provided	Yes
	Shape of Aperture provided in perforated plate	NA for wire cloth
	Thickness of perforated plate (in mm)	1
	Nominal aperture size of holes with tolerance for perforated plate (in mm)	37.5 ± 0.45
	Frame Material of test sieves	Brass sheet
	Frame material thickness (in mm)	0.45
	Depth from top edge to sieving surface (in mm)	50
	Depth from bottom edge to sieving surface (in mm)	15
	Warranty in years	1
	ISO certification of manufacturer	Yes Or higher
Availability of test certificate from Government/ NABL/ ILAC accerdated lab to prove conformity of specifications	Yes Or higher	

Additional Specification Parameters - Test Sieves As Per IS 460 (Part 1 And Part 2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional Specification Parameter 1	Test sieve, brass: 8" dia., 500µm mesh size, with stainless steel mesh
Additional Specification Parameter 2	Sieve diameter 200mm

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Pradeep Kumar Singh	403804, Director National Centre for Antarctic Ocean Research Headland Sada, Vasco-da-gama,	1	20

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**1. Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Scope of Supply

Scope of supply (Bid price to include all cost components) : Only supply of Goods

3. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

(i) The Bidder has to upload catalogue/brouchure of the offered product. (ii) The Bidders to provide basic details sought in Buyer Added Specific ATC.

4. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

5. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates

(whichever is lower), subject to the maximum of quoted GST %.

6. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

7. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

8. **Generic**

Manufacturer Authorization: Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

9. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

10. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

11. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

12. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

13. **Warranty**

Warranty period of the supplied products shall be 1 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. Any clause(s) incorporated by the Buyer regarding following shall be treated as null and void and would not be considered as part of bid:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to

exemption provided to such sellers under GeM GTC.

3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process.
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---