


राष्ट्रीय अंटार्कटिक एवं
समुद्री अनुसंधान केन्द्र
पृथ्वी विज्ञान मंत्रालय
भारत सरकार
हेड लैण्ड सडा, वास्को डा गामा
गोवा- ४०३ ८०४ भारत



**NATIONAL CENTRE FOR
ANTARCTIC & OCEAN RESEARCH**

Ministry of Earth Sciences
Government of India)
Headland Sada, Vasco da Gama
Goa - 403 804 - INDIA

Tel/Fax No.: 0091 0832 2525573

Email:warlu62@ncaor.gov.in

PROCUREMENT SECTION

Foreign Enquiry

M/s.

Ref. No.:

LAB -2487

Date:

17.02.2016

Due Date:

03.03.2016

Dear Sir,

You are requested to send your offer for the following items as per the terms and conditions mentioned below and in Annexure I.

SL. NO.	Description	Quantity
01	Water Bath (Specification as per Annexure I)	01

Your offer should contain the following information:

1. Submit your quotation on F.O.B.
2. Delivery Period:
3. Validity of quotation:
4. Quantity/Trade discounts, if any. :
5. Guarantee/Warranty of the product:
6. Taxes applicable (VAT if any):
7. Specify Brand/Make/Model of the item:
8. Enclose brochure/leaflet – Specification:
9. Mode of payment acceptable to you: No advance payment will be made. Payment will be by Telegraphic Transfer immediately after dispatch of material.

Director, NCAOR reserves the right to reject any quotation received without assigning any reasons.

Yours faithfully

**Executive (Procurement)
For and behalf of Director, NCAOR**

Specification for Water Bath.

1. Bath volume: ≤ 26 L
2. Temperature Range: 0 to 100°C
3. Setting/display resolution: 0.1°C
4. Controller Type: PID Temperature control
5. Temperature stability: $\pm 0.2^\circ\text{C}$
6. Setting/display resolution: 0.1°C
7. Audible alarms
8. It should have advanced microprocessor controller.
9. Removable bottom plate for easy cleaning.
10. Temperature indication (LED).
11. Should have high capacity of heating (≤ 2 kW).
12. Weight: 14-20 kg
13. Electrical Requirements: 230 V/50 Hz
14. System should be CE certified and ETL/UL listed.

Accessories

- Test tube racks and lift-up Makrolon covers.

Other features:

- i. It should have over temperature safety regulation.
- ii. Must be corrosion free interior.

Important:

Additional Conditions:

- The product should be ISO certified.
- The supplier must provide relevant catalogue with full details of model, photograph and all features as an evidence of our tender specifications.
- Supplier should quote only the proven model being used in reputed Govt. Research Institute involved in specialized microbial exploration.
- The supplier must provide list of users with address and phone number to whom identical type of water bath has been supplied in the last two years time.
- All safety features must be incorporative.
- Free installation.
- Supply of operating manuals required for operating the instrument.
- If quoting on behalf of foreign manufacturer an authorization letter must be enclosed otherwise the quotation will not be considered.
- Manufacturer/Supplier should install and successfully run and demonstrate the instrument.
- Minimum one year warranty after installation.