

CORRIGENDUM

GLOBAL TENDER NO: NCPOR/PS/HSS-622/GT-09 FOR SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF MICROWAVE DIGESTION SYSTEM

Revision in the technical specifications of Sr. No. 11 of the Microwave Digestion System

Sr. No	Tender specification		Revised specifications	
11	Rack	Rack must be made of PTFE-TFM. Racks should have the capacity to hold (a) minimum 7 vials of 40ml or more capacity, (b) minimum 20 vials of 15 ml or capacity, (c) minimum 27 vials of 7 ml or more capacity and (d) minimum 40 vials of 4 ml or more capacity.	Rack	Rack must be made of PTFE-TFM. Racks should have the capacity to hold (a) minimum 6 vials of 40ml or more capacity, (b) minimum 18 vials of 15 ml or more capacity, (c) minimum 27 vials of 7 ml or more capacity and (d) minimum 40 vials of 4 ml or more capacity.